

Autoclave

Protocol:

1. Check there is enough water in the bottom of the autoclave, enough to cover the bolts at the bottom.
2. Screw in the bolts one by one until finger tight.
3. Open the valve on the lid.
4. Go to the plug, turn it on and set to 40 minutes, once up to required temperature (may be steam coming out of the top), close the valve and time for 15 minutes.