

Sonication

Materials

1. Sonicator
2. Cells
3. Buffer
4. 1.5mL tubes

Protocol

1. Turn on sonicator to lowest setting (1 W)
2. Insert sonicator into 1.5mL tube with cells for 10 seconds
 - a. NOTE: cells should be without LB and with buffer
3. Repeat #2 FOUR times